



SCOPE OF ACCREDITATION TO ISO/IEC 17025:2005

Quality Calibrations, Inc.
102 E. Bay Front Rd, Deale, MD 20751
Beverly Heiberger Phone: 877-747-3883

CALIBRATION

Valid to: November 6, 2012

Certificate Number: AC-1347

I. Mechanical

Table with 5 columns: PARAMETER / EQUIPMENT, RANGE, CALIBRATION AND MEASUREMENT CAPABILITY [EXPRESSED AS UNCERTAINTY(±)], REFERENCE STANDARD OR EQUIPMENT, METHOD(S). Row 1: Pipette, 0.5 to 2.0 µL to 5 000 to 10 000 µL, 0.057 µL to 6.6 µL, Mettler-Toledo Precision Balance, SOP QACCRED and ISO 8655-2.

Notes:

- 1. Calibration and Measurement Uncertainties (Expanded Uncertainty) are based on approximately a 95% confidence interval, using a coverage of k=2
2. Capabilities are available for on-site calibration activity.
3. This laboratory's capabilities include laboratory and on-site calibration services. Since field (on-site) conditions are typically more variable than those in the laboratory, larger measurement uncertainties are expected in the field (on-site) than what is reported on the scope of accreditation.
4. L represents the volume in liters.
5. This scope is part of and must be included with the Certificate of Accreditation No.AC-1347.

Handwritten signature of Keith Greenaway

Vice President

